

實驗室認證處

Taiwan Accreditation Foundation

TAF-CNLA-FG32(2)

實驗室認證處
評鑑小組認可建議表(中文)

實驗室編號：0150

評鑑類別：延展認證

評鑑日期：2014/2/25

實驗室主管

姓名：李國琪 認可 保留領域別：校正
長度

| 校正件 | 校正方法 /最高工作標準 件 | 校正範圍 | 最小不確定度 | 認可/保留 |
|---|---|--|--|---|
| KA1001 塊規 | 自訂CP-DM-0002 /KOBA 1122M Grade K | 鋼 (0.5 to 1.18) mm (1.19 to 1.50) mm (2.0 to 10.0) mm (10.5 to 20.0) mm (20.5 to 25.0) mm (50 to 100) mm 碳化鉻 (0.5 to 1.18) mm (1.19 to 1.50) mm (2.0 to 10.0) mm (10.5 to 20.0) mm (20.5 to 25.0) mm (50 to 100) mm | 0.04 μm 0.04 μm 0.05 μm 0.07 μm 0.10 μm 0.26 μm 0.04 μm 0.04 μm 0.05 μm 0.07 μm 0.09 μm 0.25 μm | <input checked="" type="radio"/> 認可 <input type="radio"/> 保留 |
| KA1003 環規 | 自訂CP-DM-0026 /MITUTOYO LGB00001 | (11 to 200) mm | 1.6 μm | <input checked="" type="radio"/> 認可 <input type="radio"/> 保留 |
| KA1009 卡尺校正器/長 度校正儀(鋼、 陶瓷) | 自訂CP-DM-0018 /MITUTOYO LGB00001 | 鋼 (0 to 1000) mm 陶瓷 (0 to 1000) mm | 5.4 μm 5.4 μm | <input checked="" type="radio"/> 認可 <input type="radio"/> 保留 |
| KA1019 量錶校正器 | 自訂CP-DM-0019 /MITUTOYO LGB00001 | (0 to 25) mm (解析度:1 μm) (0 to 5) mm (解析度:0.1 μm) | 10 μm 2 μm | <input checked="" type="radio"/> 認可 <input type="radio"/> 保留 |
| KA2003 針盤式/數位式 卡尺 針盤式/游標式 卡尺 游標式卡尺 | 自訂CP-DM-0008 /MITUTOYO 515-556 | (0 to 600) mm (數位式解析度:0.01 mm) (0 to 600) mm (游標式解析度:0.02 mm) (0 to 600) mm (游標式解析度:0.05 mm) | 內徑0.02 mm 外徑0.02 mm 內徑0.04 mm 外徑0.04 mm 內徑0.10 mm 外徑0.10 mm | <input checked="" type="radio"/> 認可 <input type="radio"/> 保留 |
| KA2004 內徑測微器 | 自訂CP-DM-0011 /MITUTOYO 516-111 516-115 | ≤ 400 mm (解析度:0.01 mm) ≤ 400 mm (解析度:0.002 mm) ≤ 400 mm (解析度:0.001 mm) | 0.01 mm 0.004 mm 0.003 mm | <input checked="" type="radio"/> 認可 <input type="radio"/> 保留 |

| | | | | |
|------------------------|---|--|---|---|
| KA2004 | 自訂CP-DM-0012 / MITUTOYO 516-111 516-115 | | | |
| KA2005 外徑測微器 | 自訂CP-DM-0009 /MITUTOYO 516-111 516-115 | (0 to 600) mm (解析度:0.01 mm) (0 to 600) mm (解析度:0.002 mm) (0 to 600) mm (解析度:0.001 mm) | 0.01 mm 0.006 mm 0.006 mm | <input checked="" type="radio"/> 認可 <input type="radio"/> 保留 |
| KA2007 三點式內徑測微器 | 自訂CP-DM-0020 /HELIOS M15 | (11 to 200) mm (解析度:1 μm) (11 to 200) mm (解析度:2 μm) (11 to 200) mm (解析度:5 μm) | 5 μm 5 μm 6 μm | <input checked="" type="radio"/> 認可 <input type="radio"/> 保留 |
| KA2008 數位、針盤、游標式高度計 | 自訂CP-DM-0017 /MITUTOYO 515-551 | (0 to 600) mm (解析度:0.02 mm) (0 to 600) mm (解析度:0.01 mm) | 0.02 mm 0.02 mm | <input checked="" type="radio"/> 認可 <input type="radio"/> 保留 |
| KA2010 針盤指示計 | 自訂CP-DM-0010 /Mitutoyo 4143&506064 | (0 to 25) mm (解析度:0.01 mm) (0 to 5) mm (解析度:0.001 mm) | 0.02 mm 0.003 mm | <input checked="" type="radio"/> 認可 <input type="radio"/> 保留 |
| KA2012 深度校正儀 | 自訂CP-DM-0041 /MITUTOYO 516-115 | (0 to 150) mm (0 to 300) mm | 4.6 μm 4.7 μm | <input checked="" type="radio"/> 認可 <input type="radio"/> 保留 |
| KA3001 角度塊規 | 自訂CP-AN-0008 /AA GAGE R-11562-3 | 1" to 45° | 0.65" | <input checked="" type="radio"/> 認可 <input type="radio"/> 保留 |
| KA3002 直角規 | 自訂CP-AN-0011 /FUJITA 451E | (120 to 500) mm | 3.2 μm | <input checked="" type="radio"/> 認可 <input type="radio"/> 保留 |
| KA3003 方規 | 自訂CP-AN-0014 /Muller-Wedel ELCOMAT 2000 | 0° to 360° | 0.65" | <input checked="" type="radio"/> 認可 <input type="radio"/> 保留 |
| KA3004 精密分度盤 | 自訂CP-AN-0023 /Muller-Wedel ELCOMAT 2000 | 0° to 360° | 0.65" | <input checked="" type="radio"/> 認可 <input type="radio"/> 保留 |
| KA3005 多邊規 | 自訂CP-AN-0014 /Muller-Wedel ELCOMAT 2000 | 0° to 360° | 0.65" | <input checked="" type="radio"/> 認可 <input type="radio"/> 保留 |
| KA3008 自動視準儀 | 自訂CP-AN-0012 /FUJITA 700E | (-5 to +5) mm/m (解析度:0.1") (-5 to +5) mm/m (解析度:0.2") (-5 to +5) mm/m (解析度:0.5") | 0.2" 0.3" 0.7" | <input checked="" type="radio"/> 認可 <input type="radio"/> 保留 |
| KA3010 | 自訂CP-AN- | 0° to 360° | 0.1" | <input checked="" type="radio"/> 認可 |

| | | | | |
|-----------------|--|---|------------------------|---|
| 旋轉盤 | 自訂CP-AN-0027 /Muller-Wedel ELCOMAT 2000 | | | <input type="radio"/> 認可 <input type="radio"/> 保留 |
| KA3015 水平儀 | 自訂CP-AN-0007 /FUJITA 700E | (-5 to +5) mm/m (解析度: 0.02 mm/m) | 2.6" | <input checked="" type="radio"/> 認可 <input type="radio"/> 保留 |
| KA3016 電子水平儀 | 自訂CP-AN-0010 /FUJITA 700E | (-5 to +5) mm/m (解析度: 0.001 mm/m) (-5 to +5) mm/m (解析度: 0.005 mm/m) (-5 to +5) mm/m (解析度: 0.01 mm/m) | 0.2" 0.6" 1.2" | <input checked="" type="radio"/> 認可 <input type="radio"/> 保留 |
| KA3019 組合角尺 | 自訂CP-AN-0003 /HOMMEL WERKE 1602 | 0° to 90° (解析度:1') 0° to 90° (解析度:2.5') 0° to 90° (解析度:5') 0° to 90° (解析度:1°) | 1' 2.5' 5' 1° | <input checked="" type="radio"/> 認可 <input type="radio"/> 保留 |
| KA4002 平台 | 自訂CP-DM-0021 /RAHN 36IN | 方格型分割法 (900 × 500) mm 米字型分割法 (2200 × 1000) mm | 2.7 μm 4.1 μm | <input checked="" type="radio"/> 認可 <input type="radio"/> 保留 |
| KA4002 | 自訂CP-DM-0022 / | | | |
| KA4003 直規 | 自訂CP-AM-0006 /Muller-Wedel ELCOMAT 2000 | 1 m 2 m | 1.1 μm 2.2 μm | <input checked="" type="radio"/> 認可 <input type="radio"/> 保留 |

振動量/聲量

| 校正件 | 校正方法 /最高工作標準件 | 校正範圍 | 最小不確定度 | 認可/保留 |
|---------------|---|---|-------------------------|---|
| KB1001 加速規 | 自訂之加速規比較 校正程序(CP-VB-0002) /B&K 8305 | 電壓靈敏度 5 Hz to 5000 Hz 電荷靈敏度 10 Hz to 5000 Hz | 2.7 % 3.0% | <input checked="" type="radio"/> 認可 <input type="radio"/> 保留 |
| KB1002 振動計 | 自訂之振動計校正 程序書(CP-VB-0005) /B&K 8305 | 加速度:100 m/s ² 速度:500 mm/s 位移:5 mm | 3.0 % 3.4 % 3.9 % | <input checked="" type="radio"/> 認可 <input type="radio"/> 保留 |
| KB2004 噪音計 | 自訂之噪音計校正 程序(CP-SL-0001) /DELTA OHM HD 9102 | 94 dB 110 dB @1000 Hz | 1 dB | <input checked="" type="radio"/> 認可 <input type="radio"/> 保留 |

質量/力量

| 校正件 | 校正方法 /最高工作標準件 | 校正範圍 | 最小不確定度 | 認可/保留 |
|--------------|---|--|--|---|
| KC1001 法碼 | 自訂CP-MA-0002, 自訂CP-MA-0003, 自訂CP-MA-0004, 自訂CP-MA-0005, 自訂CP-MA-0009, 自訂CP-MA-0013, /Sartorius C5s Sartorius C50s Sartorius C1000s Sartorius C10000s Mettler XP26003L | 1 mg 2 mg 5 mg 10 mg 20 mg 50 mg 100 mg 200 mg 500 mg 1 g 2 g 5 g | 0.0019 mg 0.0019 mg 0.0019 mg 0.0019 mg 0.0019 mg 0.0019 mg 0.0023 mg 0.0026 mg 0.0031 mg 0.0037 mg 0.0045 mg 0.0086 mg | <input checked="" type="radio"/> 認可 <input type="radio"/> 保留 |

| | | | | |
|--------------------------|---|--|--|---|
| | | 10 g 20 g 50 g 100 g 200 g 500 g 1 kg 2 kg 5 kg 10 kg 20 kg | 0.013 mg 0.013 mg 0.019 mg 0.039 mg 0.054 mg 0.12 mg 0.23 mg 1.0 mg 1.4 mg 2.4 mg 5.0 mg | |
| KC1001 重力法碼 | 自訂CP-MA-0008 / Mettler MC210S Sartorius LA3200D Mettler XP26003L Mettler KA30-3 | (0 to 200) g 200 g to 1 kg (1 to 26) kg (26 to 30) kg | 0.17 mg 2.0 mg 17 mg 72 mg | |
| KC1002 天平 (含遊校) | 自訂CP-MA-0007 / Mettler MC210S Sartorius LA3200D Mettler XP26003L Mettler KA30-3 | (0 to 5) g (5 to 20) g (20 to 50) g (50 to 100) g (100 to 210) g 210 g to 1 kg (1 to 6) kg (6 to 15) kg (15 to 26) kg (26 to 32) kg | 0.15 mg 0.24 mg 0.29 mg 0.46 mg 1.1 mg 5.0 mg 86 mg 0.22 g 0.37 g 0.46 g | <input checked="" type="radio"/> 認可 <input type="radio"/> 保留 |
| KC1004 平台秤 (含遊校) | 自訂CP-MA-0007 / Mettler KC500-1 | (32 to 100) kg (100 to 300) kg (300 to 600) kg | 5.2 g 11 g 17 g | <input checked="" type="radio"/> 認可 <input type="radio"/> 保留 |
| KC1004 吊秤 (含遊校) | 自訂CP-MA-0013 / N.A. F1 | (0 to 600) kg(解析度: 0.1 Kg) (0 to 600) kg(解析度: 0.2 Kg) (0 to 1000) kg(解析度: 0.2 Kg) (0 to 1000) kg(解析度: 0.5 Kg) | 0.2 kg 0.4 kg 0.4 kg 1.0 kg | |
| KC2001 檢力環 | 自訂彈性測力計比較 校正程序書(CP-FO-0002) /Morehouse Series 1000 | (0.98 to 440) kN | Tension 0.12 % Compression 0.12 % of full scale | <input checked="" type="radio"/> 認可 <input type="radio"/> 保留 |
| KC2002 荷重元 | 自訂彈性測力計比較 校正程序書(CP-FO-0002) /Morehouse Series 1000 | (0.98 to 440) kN | Tension 0.16 % Compression 0.16 % of full scale | <input checked="" type="radio"/> 認可 <input type="radio"/> 保留 |
| KC2005 材料試驗機 (含遊校) | 自訂材料試驗機校正 程序書 (CP-FO-0003) /HBM Ormond/S9, BUL-FF39-OS-**-90-251 | (0.44 to 440) kN | Tension 0.16 % Compression 0.16 % of full scale | <input checked="" type="radio"/> 認可 <input type="radio"/> 保留 |
| KC3001 標準量桶 | 自訂CP-VO-0001 / Sartorius LA3200D Mettler KC500-1 | (0 to 50) mL (50 to 100) mL (100 to 250) mL (250 to 500) mL (500 to 1000) mL (1000 to 2000) mL 10 L 20 L 50 L | 0.12 mL 0.17 mL 0.4 mL 0.8 mL 1.5 mL 1.6 mL 1.4 mL 2.6 mL 3.5 mL | <input checked="" type="radio"/> 認可 <input type="radio"/> 保留 |
| KC4001 | 自訂扭力工具校正程 | (0.015 to 1000) N m | 2.0 % | <input checked="" type="radio"/> 認可 |

| | | | | |
|-----------------|--|---------------------|-------|---|
| 扭力扳手 | 序(CP-TQ-0001) /Norbar 50031.ETS | | | <input type="radio"/> 保留 |
| KC4002 扭力起子 | 自訂扭力工具校正程序(CP-TQ-0001) /Norbar 50027.ETS | (0.015 to 50) N m | 2.0 % | <input checked="" type="radio"/> 認可 <input type="radio"/> 保留 |
| KC4003 扭力計 | 自訂扭力感測器校正程序(CP-TQ-0002) /Norbar 21515 | (0.015 to 5424) N m | 0.1 % | <input checked="" type="radio"/> 認可 <input type="radio"/> 保留 |
| KC4004 扭力校正器 | 自訂扭力感測器校正程序(CP-TQ-0002) /Norbar 21515 | (0.015 to 5424) N m | 0.1 % | <input checked="" type="radio"/> 認可 <input type="radio"/> 保留 |
| KC4005 扭力倍力器 | 自訂扭力倍力器校正程序(CP-TQ-0003) /Norbar 50075.ETS | (1.0 to 4000) N m | 1.0 % | <input checked="" type="radio"/> 認可 <input type="radio"/> 保留 |

壓力量／真空量

| 校正件 | 校正方法 /最高工作標準件 | 校正範圍 | 最小不確定度 | 認可/保留 |
|----------------------------|---|--|--|---|
| KD1001 氣體活塞壓力計 | 自訂氣體活塞壓力計校正程序(CP-PR-0010) 自訂壓力計校正程序(CP-PR-0360) /RUSKA 2465 RUSKA 2400 | 1.38 kPa to 6.89 MPa 41.4 kPa to 83.7 MPa | 5.4 E-05 1.1 E-04 | <input checked="" type="radio"/> 認可 <input type="radio"/> 保留 |
| KD1002 液體活塞壓力計 | 自訂液體活塞壓力計校正程序書(CP-PR-0011) 自訂壓力計校正程序(CP-PR-0360) /RUSKA 2400 RUSKA 2485 | 41.4 kPa to 83.7 MPa 68.9 kPa to 414 MPa | 1.1 E-04 1.5 E-04 | <input checked="" type="radio"/> 認可 <input type="radio"/> 保留 |
| KD1003 汞柱壓力計 (含遊校) | 自訂汞柱壓力計校正程序技術手冊(CP-PR-0100) 自訂壓力計校正程序(CP-PR-0360) /RUSKA 2465、SCHWIEN 1025LX110 | 0 Pa to 373 kPa(RUSKA 2465) 0 Pa to 373 kPa(SCHWIEN 1025LX110) | 0.95 Pa + 2.2 E-05 1.7 Pa + 2.2 E-05 | <input checked="" type="radio"/> 認可 <input type="radio"/> 保留 |
| KD1004 壓力錶 (含遊校) | 自訂壓力計校正程序(CP-PR-0360) 自訂壓力計遊校程序(CP-PR-0361) /SCHWIEN 1025LX110 RUSKA 2465 RUSKA 2400 RUSKA 2485 | 0 Pa to 373 kPa 1.38 kPa to 6.89 MPa 41.4 kPa to 83.7 MPa 68.9 kPa to 414 MPa | 1.7 Pa + 2.2 E-5 5.4 E-05 1.1 E-04 1.5 E-04 | <input checked="" type="radio"/> 認可 <input type="radio"/> 保留 |
| KD1005 壓力轉換器 (含遊校) | 自訂壓力計校正程序(CP-PR-0360) 自訂壓力計遊校程序(CP-PR-0361) /SCHWIEN 1025LX110 RUSKA 2465 RUSKA 2400 RUSKA 2485 | 0 Pa to 373 kPa 1.38 kPa to 6.89 MPa 41.4 kPa to 83.7 MPa 68.9 kPa to 414 MPa | 1.7 Pa + 2.2 E-5 5.4 E-05 1.1 E-04 1.5 E-04 | <input checked="" type="radio"/> 認可 <input type="radio"/> 保留 |
| KD1006 空用數據測試儀 (含遊校) | 自訂6610 ADTS空用數據測試儀... 等 CP-PR-2120 /RUSKA 2465 SCHWIEN 1025LX110 RUSKA 7750i | 高度 -914.40 m to 0.00 m (-3 kft to 0 kft) 0.00 m to 9144.00 m (0 kft to 30 kft) 9144.00 m to 15240.00 m (30 kft to 50 kft) 15240.0 m to 25384.0 m (50 kft to 80 kft) 24384.0 m to 30480.0 m (80 kft to 100 kft) | 0.37 m(1.2 ft) 0.43 m(1.4 ft) 0.67 m(2.2 ft) 2.4 m(7.8 ft) 6.1 m(20 ft) 0.062 m/s(0.12 Knots) 0.024 m/s(0.046 Knots) 0.016 m/s(0.030 Knots) 0.0072 m/s(0.014 Knots) 0.0057 m/s(0.011 knots) | <input checked="" type="radio"/> 認可 <input type="radio"/> 保留 |

| | | | | |
|------------------------------------|---|--|-------------------------------|---|
| | | 空速 12.861 m/s to 33.439 m/s (25 Knots to 65 Knots) 33.439 m/s to 51.444 m/s (65 Knots to 100 Knots) 51.444 m/s to 128.6111 m/s (100 Knots to 250 Knots) 128.6111 m/s to 180.0556 m/s (250 Knots to 350 Knots) 180.0556 m/s to 514.4444 m/s (350 Knots to 1000 Knots) | | |
| KD1007 差壓計 (含遊校) | 自訂壓力計校正程序(CP-PR-0360) 自訂壓力計遊校程序(CP-PR-0361) /SCHWIEN 1025LX110 RUSKA 2465 | 0 Pa to 373 kPa 1.38 kPa to 6.89 MPa | 1.7 Pa + 2.2 E-0.5 5.4 E-5 | <input checked="" type="radio"/> 認可 <input type="radio"/> 保留 |
| KD2001 旋轉轉子黏 滯式真空計 | 自訂真空計校正程序(CP-VA-0001) /MKS SRG-2 MKS 390HA-0001 (BAZERS PSK100) | (5.00 E-05 to 1.33 E-04) Pa (>1.33 E-04 to 1.33 E-03) Pa (>1.33 E-03 to 1.33 E+01) Pa | 32 % 12 % 4.0 % | <input checked="" type="radio"/> 認可 <input type="radio"/> 保留 |
| KD2002 離子真空計 (含遊校) | 自訂真空計校正程序(CP-VA-0001) /MKS SRG-2 MKS 390HA-0001 (BAZERS PSK100) | (5.00 E-05 to 1.33 E-04) Pa (>1.33 E-04 to 1.33 E-03) Pa (>1.33 E-03 to 1.33 E-01) Pa | 33 % 13 % 7.2 % | <input checked="" type="radio"/> 認可 <input type="radio"/> 保留 |
| KD2003 電容式真空 計 (含遊校) | 自訂真空計校正程序(CP-VA-0001) 自訂單階膨脹法真空計校正程序(CP-VA-0140) /MKS 390HA-01000 MKS 390HA-00001 (MKS PVS-6B-HA) | (1.33 E-01 to 1.33 E+00) Pa (> 1.33 E+00 to 1.33 E+03) Pa (> 1.33 E+03 to 1.33 E+05) Pa | 2.7 % 1.5 % 0.32 % | <input checked="" type="radio"/> 認可 <input type="radio"/> 保留 |
| KD2004 壓縮式真空 計 (含遊校) | 自訂真空計校正程序(CP-VA-0001) /MKS 390HA-01000 MKS 390HA-00001 (MKS PVS-6B-HA) | (1.33 E-01 to 1.33 E+00) Pa (> 1.33 E+00 to 1.33 E+03) Pa | 3.5 % 1.9 % | <input checked="" type="radio"/> 認可 <input type="radio"/> 保留 |
| KD2005 熱偶真空計 (含遊校) | 自訂真空計校正程序(CP-VA-0001) /MKS 390HA-01000 MKS 390HA-00001 (MKS PVS-6B-HA) | (1.33 E-01 to 1.33 E+00) Pa (> 1.33 E+00 to 1.33 E+03) Pa (> 1.33 E+03 to 1.33 E+05) Pa | 5.0 % 4.3 % 3.6 % | <input checked="" type="radio"/> 認可 <input type="radio"/> 保留 |
| KD2006 派籃尼真空 計 數字式氣壓 計 | 自訂真空計校正程序(CP-VA-0001) /MKS 390HA-01000 MKS 390HA-00001 (MKS PVS-6B-HA) | (1.33 E-01 to 1.33 E+00) Pa (> 1.33 E+00 to 1.33 E+03) Pa (> 1.33 E+03 to 1.33 E+05) Pa | 3.5 % 1.9 % 0.34 % | <input checked="" type="radio"/> 認可 <input type="radio"/> 保留 |

| 校正件 | 校正方法 /最高工作標準件 | 校正範圍 | 最小不確定度 | 認可/保留 |
|-------------------------------------|--|---|---|---|
| KE1001 玻璃溫度計 | 自訂溫度計校正程序書(CP-TE-0010) /HART 5626 | (-50 to 250) °C | 0.09 °C | <input checked="" type="radio"/> 認可 <input type="radio"/> 保留 |
| KE1002 白金電阻溫度計 | 自訂溫度計校正程序書(CP-TE-0010) /HART 5626/HART5638 | (-50 to 250) °C (> 250 to 600) °C (>600 to 1000) °C | 0.033 °C 0.057 °C 0.83 °C | <input checked="" type="radio"/> 認可 <input type="radio"/> 保留 |
| KE1004 熱電偶 | 自訂溫度計校正程序書(CP-TE-0010) /HART 5626/ HART 5638 | (-50 to 600) °C (>600 to 1000) °C | 0.50 °C 1.0 °C | <input checked="" type="radio"/> 認可 <input type="radio"/> 保留 |
| KE1005 熱電偶溫度計 (實驗室) | 自訂溫度計校正程序書(CP-TE-0010) /HART 5626/ HART 5638 | (-50 to 600) °C (>600 to 1000) °C | 0.50 °C 1.0 °C | <input checked="" type="radio"/> 認可 <input type="radio"/> 保留 |
| KE1006 溫度指示錶 | 自訂溫度模擬校正程序書(CP-TE-0004) /EDC 522 | J TYPE (-50 to 1000) °C K TYPE (-50 to 1000) °C T TYPE (-50 to 400) °C E TYPE (-50 to 1000) °C R TYPE (0 to 1000) °C S TYPE (0 to 1000) °C RTD (0 to 600) °C | 0.49 °C 0.49 °C 0.38 °C 0.49 °C 0.38 °C 0.38 °C 0.38 °C 0.1 °C | <input checked="" type="radio"/> 認可 <input type="radio"/> 保留 |
| KE1008 熱電偶模擬器 | 自訂溫度模擬校正程序書(CP-TE-0004) /EDC 522 | J TYPE (-50 to 1000) °C K TYPE (-50 to 1000) °C T TYPE (-50 to 400) °C E TYPE (-50 to 1000) °C R TYPE (0 to 1000) °C S TYPE (0 to 1000) °C RTD (-50 to 600) °C | 0.49 °C 0.49 °C 0.38 °C 0.49 °C 0.38 °C 0.38 °C 0.1 °C | <input checked="" type="radio"/> 認可 <input type="radio"/> 保留 |
| KE1009 溫度校正器 | 自訂溫度計校正程序書(CP-TE-0010) /HART 5626 HART 5638 | (-50 to 600) °C (>600 to 1000) °C | 0.50 °C 1.0 °C | <input checked="" type="radio"/> 認可 <input type="radio"/> 保留 |
| KE1010 溫度櫃,高 溫爐,烘箱, 烤箱(遊校) | 自訂溫度遊校程序書(CP-TE-0005) /FLUKE 724 | (-50 to 1000) °C | 1.2 °C | <input checked="" type="radio"/> 認可 <input type="radio"/> 保留 |
| KE1011 熱電偶溫度 | 自訂之溫度遊校程序書(CP-TE- | (-50 to 1000) °C (-50 to 1000) °C | 1.2 °C 0.9 °C | <input checked="" type="radio"/> 認可 <input type="radio"/> 保留 |

| | | | | |
|--------------------------------|--|--------------------------------|-------------------|---|
| 計, 溫度指示錶(溫度櫃, 高溫爐, 烘箱, 烤箱)(遊校) | 0005) /FLUKE 724 | | | |
| KE2001 濕度計 | 自訂濕度校正程序(CP-HU-0001) /THUNDER 2500ST | (10 to 90) %RH (5 to 60) °C | 1.0 %RH 0.2 °C | <input checked="" type="radio"/> 認可 <input type="radio"/> 保留 |
| KE2003 露點計 | 自訂露點計校正程序(CP-HU-0002) /RH SYSTEMS 373L | (-60 to 0) °C | 0.9 °C | <input checked="" type="radio"/> 認可 <input type="radio"/> 保留 |
| KE2004 溫濕度記錄器 | 自訂濕度校正程序(CP-HU-0001) /THUNDER 2500ST | (10 to 90) %RH (5 to 60) °C | 1.2 %RH 0.7 °C | <input checked="" type="radio"/> 認可 <input type="radio"/> 保留 |

電量

| 校正件 | 校正方法 /最高工作標準件 | 校正範圍 | 最小不確定度 | 認可/保留 |
|-----------------------------------|--|--|---|---|
| KF1001 直流電壓源 直流電壓源 直流電壓錶 | 自訂直流電壓校正程序書(CP-SD-V001)、 自訂多功能校正器校正程序書(CP-SD-0001)、 自訂多功能電表校正程序書(CP-MT-0001) /DATRON 4910、 FLUKE 5700A、 DATRON 1281 | 1 mV 10 mV 0.1 V 1.0 V 1.018 V 10 V 100 V 1000 V | 1.1 mV/V 0.13 mV/V 7.8 µV/V 1.3 µV/V 1.3 µV/V 1.2 µV/V 4.5 µV/V 5.6 µV/V | <input checked="" type="radio"/> 認可 <input type="radio"/> 保留 |
| KF1001 | 自訂多功能校正器線性校正程序書(CP-SD-0004) /FLUKE 5700A | < 100 mV 0.1 V to 1.0 V 1.0 V to 10 V 10 V to 100 V 100 V to 1000V | 32 µV/V 17 µV/V 17 µV/V 21 µV/V 23 µV/V | |
| KF1002 直流電流源 直流電流錶 | 自訂多功能校正器校正程序書(CP-SD-0001)、 自訂多功能電表校正程序書(CP-MT-0001) /FLUKE 5700A、 DATRON 1281 | 100 µA 1 mA 10 mA 100 mA 1 A | 71 µA/A 44 µA/A 46 µA/A 60 µA/A 67 µA/A | <input checked="" type="radio"/> 認可 <input type="radio"/> 保留 |
| KF1002 | 自訂多功能校正器線性校正程序書(CP-SD-0004) /FLUKE 5700A | < 100 µA 0.1 mA to 1.0 mA 1.0 mA to 10 mA 10 mA to 100 mA 100 mA to 1000 mA | 94 µA/A 52 µA/A 54 µA/A 67 µA/A 96 µA/A | |
| KF1011 交直流熱轉換器 | 自訂交直流熱轉換器校正程序書(CP-SD-0002) /FLUKE 792A | 2 mV(10 Hz) 2 mV(20 Hz) 2 mV(40 Hz) 2 mV(100 Hz) 2 mV(1 kHz) 2 mV(10 kHz) 2 mV(20 kHz) 2 mV(50 kHz) | 2.8 mV/V 2.1 mV/V 2.0 mV/V 2.1 mV/V 1.9 mV/V 2.0 mV/V 2.2 mV/V 1.9 mV/V | <input checked="" type="radio"/> 認可 <input type="radio"/> 保留 |

| | | | | |
|--------|---|--|---|--|
| | | 2 mV(100 kHz) 2 mV(300 kHz) 2 mV(500 kHz) 2 mV(1 MHz) | 2.5 mV/V 3.2 mV/V 5.6 mV/V 6.5 mV/V | |
| KF1011 | / | 10 mV(10 Hz) 10 mV(20 Hz) 10 mV(40 Hz) 10 mV(100 Hz) 10 mV(1 kHz) 10 mV(10 kHz) 10 mV(20 kHz) 10 mV(50 kHz) 10 mV(100 kHz) 10 mV(300 kHz) 10 mV(500 kHz) 10 mV(1 MHz) | 0.39 mV/V 0.39 mV/V 0.34 mV/V 0.36 mV/V 0.40 mV/V 0.36 mV/V 0.36 mV/V 0.47 mV/V 0.67 mV/V 0.85 mV/V 1.6 mV/V 1.8 mV/V | |
| KF1011 | / | 20 mV(10 Hz) 20 mV(20 Hz) 20 mV(40 Hz) 20 mV(100 Hz) 20 mV(1 kHz) 20 mV(10 kHz) 20 mV(20 kHz) 20 mV(50 kHz) 20 mV(100 kHz) 20 mV(300 kHz) 20 mV(500 kHz) 20 mV(1 MHz) | 0.34 mV/V 0.22 mV/V 0.18 mV/V 0.19 mV/V 0.18 mV/V 0.18 mV/V 0.18 mV/V 0.28 mV/V 0.58 mV/V 0.83 mV/V 1.4 mV/V 1.6 mV/V | |
| KF1011 | / | 100 mV(10 Hz) 100 mV(20 Hz) 100 mV(40 Hz) 100 mV(100 Hz) 100 mV(1 kHz) 100 mV(10 kHz) 100 mV(20 kHz) 100 mV(50 kHz) 100 mV(100 kHz) 100 mV(300 kHz) 100 mV(500 kHz) 100 mV(1 MHz) | 0.25 mV/V 0.13 mV/V 0.060 mV/V 0.070 mV/V 0.067 mV/V 0.065 mV/V 0.065 mV/V 0.13 mV/V 0.19 mV/V 0.54 mV/V 0.72 mV/V 0.74 mV/V | |
| KF1011 | / | 200 mV(10 Hz) 200 mV(20 Hz) 200 mV(40 Hz) 200 mV(100 Hz) 200 mV(1 kHz) 200 mV(10 kHz) 200 mV(20 kHz) 200 mV(50 kHz) 200 mV(100 kHz) 200 mV(300 kHz) 200 mV(500 kHz) 200 mV(1 MHz) | 0.26 mV/V 0.11 mV/V 0.064 mV/V 0.065 mV/V 0.064 mV/V 0.061 mV/V 0.065 mV/V 0.13 mV/V 0.19 mV/V 0.54 mV/V 0.71 mV/V 0.73 mV/V | |
| KF1011 | / | 600 mV(10 Hz) 600 mV(20 Hz) 600 mV(40 Hz) 600 mV(100 Hz) 600 mV(1 kHz) 600 mV(10 kHz) | 0.24 mV/V 0.11 mV/V 0.043 mV/V 0.032 mV/V 0.033 mV/V 0.032 mV/V | |

| | | | | |
|--------|---|--|--|--|
| | | 600 mV(20 kHz) 600 mV(50 kHz) 600 mV(100 kHz) 600 mV(300 kHz) 600 mV(500 kHz) 600 mV(1 MHz) | 0.030 mV/V 0.061 mV/V 0.082 mV/V 0.18 mV/V 0.51 mV/V 0.65 mV/V | |
| KF1011 | / | 1 V (10 Hz) 1 V(20 Hz) 1 V(40 Hz) 1 V(100 Hz) 1 V(1 kHz) 1 V(10 kHz) 1 V(20 kHz) 1 V(50 kHz) 1 V(100 kHz) 1 V(300 kHz) 1 V(500 kHz) 1 V(1 MHz) | 0.25 mV/V 0.078 mV/V 0.037 mV/V 0.019 mV/V 0.020 mV/V 0.020 mV/V 0.020 mV/V 0.055 mV/V 0.068 mV/V 0.16 mV/V 0.51 mV/V 0.53 mV/V | |
| KF1011 | / | 2 V(10 Hz) 2 V(20 Hz) 2 V(40 Hz) 2 V(100 Hz) 2 V(1 kHz) 2 V(10 kHz) 2 V(20 kHz) 2 V(50 kHz) 2 V(100 kHz) 2 V(300 kHz) 2 V(500 kHz) 2 V(1 MHz) | 0.24 mV/V 0.084 mV/V 0.043 mV/V 0.034 mV/V 0.032 mV/V 0.032 mV/V 0.032 mV/V 0.063 mV/V 0.074 mV/V 0.17 mV/V 0.53 mV/V 0.56 mV/V | |
| KF1011 | / | 6 V(10 Hz) 6 V(20 Hz) 6 V(40 Hz) 6 V(100 Hz) 6 V(1 kHz) 6 V(10 kHz) 6 V(20 kHz) 6 V(50 kHz) 6 V(100 kHz) 6 V(300 kHz) 6 V(500 kHz) 6 V(1 MHz) | 0.25 mV/V 0.083 mV/V 0.044 mV/V 0.030 mV/V 0.032 mV/V 0.034 mV/V 0.032 mV/V 0.055 mV/V 0.074 mV/V 0.17 mV/V 0.52 mV/V 0.57 mV/V | |
| KF1011 | / | 10 V(10 Hz) 10 V(20 Hz) 10 V(40 Hz) 10 V(100 Hz) 10 V(1 kHz) 10 V(10 kHz) 10 V(20 kHz) 10 V(50 kHz) 10 V(100 kHz) 10 V(300 kHz) 10 V(500 kHz) 10 V(1 MHz) | 0.23 mV/V 0.081 mV/V 0.037 mV/V 0.025 mV/V 0.025 mV/V 0.024 mV/V 0.025 mV/V 0.054 mV/V 0.069 mV/V 0.16 mV/V 0.51 mV/V 0.57 mV/V | |
| KF1011 | / | 20 V(10 Hz) 20 V(20 Hz) 20 V(40 Hz) 20 V(100 Hz) | 0.24 mV/V 0.085 mV/V 0.042 mV/V 0.024 mV/V | |

| | | | | |
|--------|---|--|--|---|
| | | 20 V(100 Hz) 20 V(1 kHz) 20 V(10 kHz) 20 V(20 kHz) 20 V(50 kHz) 20 V(100 kHz) 20 V(300 kHz) 20 V(500 kHz) 20 V(1 MHz) | 0.034 mV/V 0.033 mV/V 0.032 mV/V 0.034 mV/V 0.073 mV/V 0.096 mV/V 0.20 mV/V 0.51 mV/V 0.57 mV/V | |
| KF1011 | / | 60 V(10 Hz) 60 V(20 Hz) 60 V(40 Hz) 60 V(100 Hz) 60 V(1 kHz) 60 V(10 kHz) 60 V(20 kHz) 60 V(50 kHz) 60 V(100 kHz) 60 V(300 kHz) | 0.25 mV/V 0.086 mV/V 0.053 mV/V 0.052 mV/V 0.050 mV/V 0.051 mV/V 0.055 mV/V 0.085 mV/V 0.097 mV/V 0.17 mV/V | |
| KF1011 | / | 100 V(10 Hz) 100 V(20 Hz) 100 V(40 Hz) 100 V(100 Hz) 100 V(1 kHz) 100 V(10 kHz) 100 V(20 kHz) 100 V(50 kHz) 100 V(100 kHz) | 0.26 mV/V 0.080 mV/V 0.043 mV/V 0.040 mV/V 0.038 mV/V 0.038 mV/V 0.038 mV/V 0.085 mV/V 0.092 mV/V | |
| KF1011 | / | 200 V(10 Hz) 200 V(20 Hz) 200 V(40 Hz) 200 V(100 Hz) 200 V(1 kHz) 200 V(10 kHz) 200 V(20 kHz) 200 V(50 kHz) 200 V(100 kHz) | 0.27 mV/V 0.12 mV/V 0.060 mV/V 0.051 mV/V 0.049 mV/V 0.052 mV/V 0.055 mV/V 0.084 mV/V 0.11 mV/V | |
| KF1011 | / | 1000 V(40 Hz) 1000 V(100 Hz) 1000 V(1 kHz) 1000 V(10 kHz) 1000 V(20 kHz) | 0.037 mV/V 0.036 mV/V 0.037 mV/V 0.038 mV/V 0.042 mV/V | |
| KF1011 | 自訂多功能校正器校 交流電 壓源 交流電 壓錶 自訂多功能電表校正 程序書(CP-MT- 0001) /FLUKE 5700A、 DATRON 1281 | 0.1 V 1.0 V 10 V 100 V 1000 V (1 kHz) 0.1 V 1.0 V 10 V 100 V 1000 V (60 Hz) | 0.25 mV/V 0.092 mV/V 0.092 mV/V 0.12 mV/V 0.13 mV/V 0.25 mV/V 0.092 mV/V 0.092 mV/V 0.12 mV/V 0.13 mV/V | <input checked="" type="radio"/> 認可 <input type="radio"/> 保留 |
| KF1011 | 自訂多功能校正器線 性校正程序書(CP- SD-0004) /FLUKE 5700A | < 100 mV 0.1 V to 1.0 V 1.0 V to 10 V 10 V to 100 V 100 V to 1000 V | 0.26 mV/V 0.11 mV/V 0.11 mV/V 0.13 mV/V 0.15 mV/V | |

| | | 100 V TO 1000 V (1 kHz) | 0.15 mV/V | |
|--------|--|---|---|---|
| KF1012 | 自訂多功能校正器校 交流電 流源 交流電 流錶 自訂多功能電錶校正 程序(CP-MT-0001) /FLUKE 5700A、 DATRON-1281 | 100 μ A 1 mA 10 mA 100 mA 1 A (1 kHz) 100 μ A 1 mA 10 mA 100 mA 1 A (60 Hz) | 0.45 mA/A 0.20 mA/A 0.20 mA/A 0.22 mA/A 0.88 mA/A 0.45 mA/A 0.20 mA/A 0.20 mA/A 0.22 mA/A 0.88 mA/A | <input checked="" type="radio"/> 認可 <input type="radio"/> 保留 |
| KF1012 | 自訂多功能校正器線 性校正程序書(CP- SD-0004) / FLUKE 5700A | ACA at 1 kHz < 100 μ A 0.1 mA to 1.0 mA 1.0 mA to 10 mA 10 mA to 100 mA 100 mA to 1000 mA | 0.56 mA/A 0.23 mA/A 0.23 mA/A 0.25 mA/A 0.92 mA/A | |
| KF3001 | 自訂直流電阻量測校 正程序書(CP-SD- R001) /ESI SR104, L&N4210B | 0.001 Ω 0.01 Ω 0.1 Ω 1 Ω 10 Ω 100 Ω 1 k Ω 10 k Ω 100 k Ω 1 M Ω | 68 $\mu\Omega/\Omega$ 50 $\mu\Omega/\Omega$ 38 $\mu\Omega/\Omega$ 2.4 $\mu\Omega/\Omega$ 6.4 $\mu\Omega/\Omega$ 8.8 $\mu\Omega/\Omega$ 6.0 $\mu\Omega/\Omega$ 1.2 $\mu\Omega/\Omega$ 12 $\mu\Omega/\Omega$ 17 $\mu\Omega/\Omega$ | <input checked="" type="radio"/> 認可 <input type="radio"/> 保留 |
| KF3001 | 自訂多功能校正器校 正程序書(CP-SD- 0001) 自訂多功能電表校正 程序書(CP-MT- 0001) / FLUKE-5700A DATRON-1281 | 1 Ω 10 Ω 100 Ω 1 k Ω 10 k Ω 100 k Ω 1 M Ω 10 M Ω | 0.14 m Ω/Ω 41 $\mu\Omega/\Omega$ 27 $\mu\Omega/\Omega$ 20 $\mu\Omega/\Omega$ 18 $\mu\Omega/\Omega$ 30 $\mu\Omega/\Omega$ 31 $\mu\Omega/\Omega$ 58 $\mu\Omega/\Omega$ | |
| KF3001 | 多功能校正器線性校 正程序書(CP-SD- 0004) / FLUKE-5700A | 1 Ω to 10 Ω 10 Ω to 100 Ω 100 Ω to 1 k Ω 1 k Ω to 10 k Ω 10 k Ω to 100 k Ω 100 k Ω to 1 M Ω 1 M Ω to 10 M Ω | 0.15 m Ω/Ω 46 $\mu\Omega/\Omega$ 50 $\mu\Omega/\Omega$ 45 $\mu\Omega/\Omega$ 49 $\mu\Omega/\Omega$ 34 $\mu\Omega/\Omega$ 0.28 m Ω/Ω | |
| KF3002 | 自訂標準電感量測校 正程序書(CP-SD- L001) /GR-1689M | 100 μ H 1 mH 10 mH 100 mH 1 H 10 H (100 Hz) 100 μ H 1 mH 10 mH 100 mH 1 H 10 H (1 kHz) | 3.0 mH/H 0.6 mH/H 0.6 mH/H 0.6 mH/H 0.6 mH/H 0.6 mH/H (100 Hz) (100 Hz) 3.0 mH/H 0.5 mH/H 0.6 mH/H 0.6 mH/H 2.0 mH/H 5.0 mH/H (1 kHz) | <input checked="" type="radio"/> 認可 <input type="radio"/> 保留 |

| | | | | |
|-----------------------------|--|---------|-----------|---|
| KF3003 電容器 電容錶/ 電橋 | 自訂標準電容量測校正程序書(CP-SD-C001) /AH-2500A | 1 pF | 7.0 μF/F | <input checked="" type="radio"/> 認可 <input type="radio"/> 保留 |
| | | 10 pF | 6.0 μF/F | |
| | | 100 pF | 6.0 μF/F | |
| | | 1000 pF | 13 μF/F | |
| | | 0.01 μF | 0.15 mF/F | |
| | | 0.1 μF | 0.15 mF/F | |
| | | 1 μF | 0.15 mF/F | |
| | | (1 kHz) | (1 kHz) | |

電磁量

| 校正件 | 校正方法 /最高工作標準件 | 校正範圍 | 最小不確定度 | 認可/保留 |
|---|--|--------------------------|----------------|---|
| KG1002 衰減器,終端器,短路器,開路器,不匹配器,微波元件,向量網路分析儀 | 自訂微波散射參數校正程序書(CP-SD-1702) /HP 8510C | 頻率: 45 Mhz to 26.5 GHz | 反射係數: | <input checked="" type="radio"/> 認可 <input type="radio"/> 保留 |
| | | 1. 反射係數: 0 to 1 | | |
| | | 2. 透射係數: 10 dB to -60 dB | | |
| | | 反射係數: | | |
| | | 45 MHz to 2 GHz | 0.0062 (1.17°) | |
| | | 3.5 mm Type | 0.0062 (0.93°) | |
| | | N Type | 0.013 (1.63°) | |
| | | 7 mm Type | | |
| | | > 2 GHz to 8 GHz | 0.0085 (3.30°) | |
| | | 3.5 mm Type | 0.011 (3.17°) | |
| | | N Type | 0.013 (3.50°) | |
| | | 7 mm Type | | |
| | | > 8 GHz to 20 GHz | 0.0086 (5.78°) | |
| 3.5 mm Type | | | | |
| > 8 GHz to 18 GHz | 0.011 (5.53°) | | | |
| N Type | 0.013 (5.98°) | | | |
| 7 mm Type | | | | |
| > 20 GHz to 26.5 GHz | 0.0087 (7.05°) | | | |
| 3.5 mm Type | | | | |
| 透射係數: | | | | |
| 45 MHz to 2 GHz | 0.054 dB (0.21°) | | | |
| 3.5 mm Type | 0.051 dB (0.21°) | | | |
| N Type | 0.059 dB (0.25°) | | | |
| 7 mm Type | | | | |
| > 2 GHz to 8 GHz | 0.064 dB (1.80°) | | | |
| 3.5 mm Type | 0.061 dB (1.79°) | | | |
| N Type | 0.061 dB (1.78°) | | | |
| 7 mm Type | | | | |
| > 8 GHz to 20 GHz | 0.094 dB (4.29°) | | | |
| 3.5 mm Type | | | | |
| > 8 GHz to 18 GHz | 0.092 dB (3.91°) | | | |
| N Type | 0.085 dB (3.84°) | | | |
| 7 mm Type | | | | |
| > 20 GHz to 26.5 GHz | 0.11 dB (5.64°) | | | |
| 3.5 mm Type | | | | |

流量

| 校正件 | 校正方法 /最高工作標準件 | 校正範圍 | 最小不確定度 | 認可/保留 |
|----------------------|--------------------------|-----------------------------------|--------|---|
| KH1001 熱質式 流量計 | 自訂鐘型氣體流量校正程序(CP-FL-0004) | (3.6 to 300) dm ³ /min | 0.45 % | <input checked="" type="radio"/> 認可 <input type="radio"/> 保留 |

| | | | | |
|---|--|------------------|--------|---|
| 浮沉式 流量計 渦輪式 流量計 | /Brooks Bell Prover 1092A | | | |
| KH1002 渦輪式 流量計 浮沉式 流量計 差壓式 流量計 孔口板 流量計 科氏力 式流量 計 | 自訂水流率校正程序 (CP-FL-0002) /EG&G OT4002T1S | (0.4 to 250) GPM | 0.16 % | <input checked="" type="radio"/> 認可 <input type="radio"/> 保留 |
| KH1003 渦輪式 流量計 浮沉式 流量計 差壓式 流量計 孔口板 流量計 科氏力 式流量 計 | 自訂油流率校正程序 (CP-FL-0001) /EG&G MT501T1A | (0.05 to 50) GPM | 0.23 % | <input checked="" type="radio"/> 認可 <input type="radio"/> 保留 |

化學量

| 校正件 | 校正方法 /最高工作標準件 | 校正範圍 | 最小不確定度 | 認可/保留 |
|---------------|---|--------------------------------|--------|---|
| KI3000 比重計 | 自訂比重計校正程 序書(CP-CM- 0001) /Anton Pear DMA 58 | 0.7990 to 1.5160 (@ 20 °C) | 0.0004 | <input checked="" type="radio"/> 認可 <input type="radio"/> 保留 |

時頻

| 校正件 | 校正方法 /最高工作標 準件 | 校正範圍 | 最小不確定度 | 認可/保留 |
|---------------------------|-----------------------------------|--------------------------|------------|---|
| KJ0100 計時器 | 自訂CP- TM-0001 /HP5071A | (0 to 900) s | 5.0 E-02 s | <input checked="" type="radio"/> 認可 <input type="radio"/> 保留 |
| KJ0200 頻率標準、頻率 信號源等 | 自訂CP-FN- 0035 /HP5071A | 1 MHz 5 MHz 10 MHz | 8.2 E-11 | <input checked="" type="radio"/> 認可 <input type="radio"/> 保留 |
| KJ0200 計頻器 | 自訂CP-FN- F002 / HP5071A | 100 KHz to 1.3 GHz | 6.4 E-09 | |
| KJ0300 轉速計 | 自訂CP-TA- 0001 /TICO 8730 | (20 to 15000) rpm | 2 rpm | <input checked="" type="radio"/> 認可 <input type="radio"/> 保留 |

報告簽署人

| 簽署人姓名 | 簽署範圍 |
|-------|--|
| 何傑 | KA1001,KA1003,KA1009,KA1019,KA2003,KA2004,KA2005,KA2007,KA2008,KA2010, KA2012,KA3001,KA3002,KA3003,KA3004,KA3005,KA3008,KA3010,KA3015,KA3016, |

| | |
|-----|---|
| | KA3019,KA4002,KA4003 |
| 劉有生 | KG1002,KJ0100,KJ0200,KJ0300 |
| 夏少康 | KF1001,KF1002,KF1011,KF1011,KF1012,KF3001,KF3002,KF3003 |
| 李國琪 | KB1001,KB1002,KB2004,KC4001,KC4002,KC4003,KC4004,KC4005,KE1001,KE1002,KE1004,KE1005,KE1006,KE1008,KE1009,KE1010,KE1011,KE2001,KE2003,KE2004 |
| 李鈺泰 | KJ0100,KJ0200,KJ0300 |
| 楊志穎 | KF1001,KF1002,KF1011,KF1011,KF1012,KF3001,KF3002,KF3003 |
| 羅允劭 | KC4001,KC4002,KC4003,KC4004,KC4005,KH1001,KH1002,KH1003 |
| 莊朝章 | KC1001,KC1002,KC1004,KC2001,KC2002,KC2005,KC3001,KE2001,KE2003,KE2004,KH1001,KH1002,KH1003,KI3000 |
| 葉東明 | KG1002,KJ0100,KJ0200,KJ0300 |
| 蔡源興 | KA1001,KA1003,KA1009,KA1019,KA2005,KA2007,KC1001,KC1002,KC1004,KC3001 |
| 陳中平 | KA2003,KA2004,KA2008,KA2010,KA2012 |
| 黃朝鈺 | KB1001,KB1002,KB2004,KC1001,KC1002,KC1004,KC3001,KD1001,KD1002,KD1003,KD1004,KD1005,KD1006,KD1007,KD2001,KD2002,KD2003,KD2004,KD2005,KD2006 |

實驗室認證處
評鑑小組認可建議表(英文)

Lab. No. : 0150

Applied Category : Renewal accreditation

Date : 2014/2/25

Laboratory head

Name : LEE, Kuo-Chi**Accreditation Field : Calibration**

Length

| Calibration items | Calibration method /Working Standard | Ranges | Smallest uncertainty |
|--|--------------------------------------|---|--|
| KA1001 gauge block | CP-DM-0002 /KOBA 1122M Grade K | Steel (0.5 to 1.18) mm (1.19 to 1.50) mm (2.0 to 10.0) mm (10.5 to 20.0) mm (20.5 to 25.0) mm (50 to 100) mm Chromium Carbide (0.5 to 1.18) mm (1.19 to 1.50) mm (2.0 to 10.0) mm (10.5 to 20.0) mm (20.5 to 25.0) mm (50 to 100) mm | 0.04 μm 0.04 μm 0.05 μm 0.07 μm 0.10 μm 0.26 μm 0.04 μm 0.04 μm 0.05 μm 0.07 μm 0.09 μm 0.25 μm |
| KA1003 Ring Gauge | CP-DM-0026 /MITUTOYO LG B00001 | (11 to 200) mm | 1.6 μm |
| KA1009 Caliper Checker/ Check Master(Steel, Ceramic) | CP-DM-0018 /MITUTOYO LG B00001 | steel (0 to 1000) mm ceramic (0 to 1000) mm | 5.4 μm 5.4 μm |
| KA1019 Dial Gauge Checker | CP-DM-0019 /MITUTOYO LG B00001 | (0 to 25) mm (resolution:1 μm) (0 to 5) mm (resolution:0.1 μm) | 10 μm 2 μm |

| | | | |
|------------|---------------|------------------------------------|-------------------|
| ker | | | |
| KA2003 | CP-DM-0008 | (0 to 600) mm | Inside : 0.02 mm |
| Dial/Digi | /MITUTOYO 51 | (Digimatic Caliper | Outside 0.02 mm |
| matic Ca | 5-556 | resolution:0.01 mm) | |
| liper | | | Inside : 0.04 mm |
| Dial/Vern | | (0 to 600) mm | Outside : 0.04 mm |
| ier Calipe | | (Vernier Caliper | |
| r | | resolution:0.02 mm) | Inside : 0.10 mm |
| Vernier C | | | Outside : 0.10 mm |
| aliper | | (0 to 600) mm(Vernier Caliper | |
| | | resolution:0.05 mm) | |
| KA2004 | CP-DM-0011 | ≤ 400 mm(resolution:0.01 mm) | 0.01 mm |
| Inside Mi | /MITUTOYO 51 | ≤ 400 mm(resolution:0.002 mm) | 0.004 mm |
| cometer | 6-111 | ≤400 mm(resolution:0.001 mm) | 0.003 mm |
| | 516-115 | | |
| KA2004 | CP-DM-0012 | | |
| | / | | |
| | MITUTOYO 51 | | |
| | 6-111 | | |
| | 516-115 | | |
| KA2005 | CP-DM-0009 | (0 to 600) mm | 0.01 mm |
| Outside | /MITUTOYO 51 | (resolution:0.01 mm) | |
| Micromet | 6-111 516-115 | (0 to 600) mm(resolution:0.002 mm) | 0.006 mm |
| er | | (0 to 600) mm | |
| | | (resolution:0.001 mm) | 0.006 mm |
| KA2007 | CP-DM-0020 | (11 to 200) mm | 5 μm |
| Three Po | /HELIOS M15 | (resolution:1 μm) | |
| ints Insid | | (11 to 200) mm | 5 μm |
| e Microm | | (resolution:2 μm) | |
| eter | | (11 to 200) mm | 6 μm |
| | | (resolution:5 μm) | |
| KA2008 | CP-DM-0017 | (0 to 600) mm | 0.02 mm |
| Digital H | /MITUTOYO 51 | (resolution:0.02 mm) | |
| eight Ga | 5-551 | 0 to 600) mm | 0.02 mm |
| uge | | (resolution:0.01 mm) | |
| Dial Heig | | | |
| ht Gauge | | | |
| Vernier H | | | |
| eight Ga | | | |
| uge | | | |
| KA2010 | CP-DM-0010 | (0 to 25) mm | 0.02 mm |
| Dial Indic | /Mitutoyo 414 | (resolution:0.01 mm) | |
| ator | 3&506064 | (0 to 5) mm | 0.003 mm |
| | | (resolution:0.001 mm) | |
| KA2012 | CP-DM-0041 | (0 to 150) mm | 4.6 μm |
| Depth C | /MITUTOYO 51 | (0 to 300) mm | 4.7 μm |
| hecker | 6-115 | | |
| KA3001 | CP-AN-0008 | 1" to 45° | 0.65" |
| Angle Ga | /AA GAGE R-11 | | |
| uge Bloc | 562-3 | | |
| k | | | |
| KA3002 | CP-AN-0011 | (120 to 500) mm | 3.2 μm |
| Square | /FUJITA 451E | | |
| KA3003 | CP-AN-0014 | 0° to 360° | 0.65" |
| True Squ | /Muller-Wedel | | |
| are | ELCOMAT 2000 | | |
| KA3004 | CP-AN-0023 | 0° to 360° | 0.65" |
| Precision | /Muller-Wedel | | |
| Index | ELCOMAT 2000 | | |
| KA3005 | CP-AN-0014 | 0° to 360° | 0.65" |
| Polygon | /Muller-Wedel | | |

| | | | |
|---------------------------------------|---|---|------------------------|
| | ELCOMAT 2000 | | |
| KA3008 Autocolli mator | CP-AN-0012 /FUJITA 700E | (-5 to +5) mm/m (resolution:0.1") (-5 to +5) mm/m (resolution:0.2") (-5 to +5) mm/m (resolution:0.5") | 0.2" 0.3" 0.7" |
| KA3010 Rotary T able | CP-AN-0027 /Muller-Wedel ELCOMAT 2000 | 0° to 360° | 2.1" |
| KA3015 Spirit Le vel | CP-AN-0007 /FUJITA 700E | (-5 to +5) mm/m (resolution:0.02 mm/m) | 2.6" |
| KA3016 Electroni c Level | CP-AN-0010 /FUJITA 700E | (-5 to +5) mm/m (resolution:0.001 mm/m) (-5 to +5) mm/m (resolution:0.005 mm/m) (-5 to +5) mm/m (resolution:0.01 mm/m) | 0.2" 0.6" 1.2" |
| KA3019 Combina tion Squ ares | CP-AN-0003 /HOMMEL WER KE 1602 | 0° to 90° (resolution:1') 0° to 90° (resolution:2.5') 0° to 90° (resolution:5') 0° to 90° (resolution:1°) | 1' 2.5' 5' 1° |
| KA4002 Surface Plate | CP-DM-0021 /RAHN 36IN | Rectangle grid method (900 × 500) m m Union jack method (2200 × 1000) m m | 2.7 μm 4.1 μm |
| KA4002 | CP-DM-0022 / | | |
| KA4003 Straight Edge | CP-AM-0006 /Muller-Wedel ELCOMAT 2000 | 1 m 2 m | 1.1 μm 2.2 μm |

Vibration & Acoustics

| Calibrati on items | Calibration met hod /Working Stan dard | Ranges | Smallest uncertainty |
|-------------------------------------|---|---|-------------------------|
| KB1001 Acceler ometer | CP-VB-0002 /B&K 8305 | 5 Hz to 5000 Hz 10 Hz to 5000 Hz | 2.7 % 3.0% |
| KB1002 Vibration Meter | CP-VB-0005 /B&K 8305 | 100 m/s ² (acceleration) 500 mm/s(velocity) 5 mm(displacement) | 3.0 % 3.4 % 3.9 % |
| KB2004 Sound L evel Met er | CP-SL-0001 /DELTA OHM H D 9102 | 94 dB 110 dB @1000 Hz | 1 dB |

Mass/Force

| Calibrati on items | Calibration met hod /Working Stan dard | Ranges | Smallest uncertainty |
|--------------------------|--|--|--|
| KC1001 Weight | CP-MA-0002, CP-MA-0003, CP-MA-0004, CP-MA-0005, CP-MA-0009, CP-MA-0013, /Sartorius C5s Sartorius C50s Sartorius C100 | 1 mg 2 mg 5 mg 10 mg 20 mg 50 mg 100 mg 200 mg 500 mg 1 g | 0.0019 mg 0.0019 mg 0.0019 mg 0.0019 mg 0.0019 mg 0.0019 mg 0.0023 mg 0.0026 mg 0.0031 mg 0.0037 mg |

| | | | |
|--|--|--|--|
| | 0s Sartorius C100 00s Mettler XP260 03L | 2 g 5 g 10 g 20 g 50 g 100 g 200 g 500 g 1 kg 2 kg 5 kg 10 kg 20 kg | 0.0045 mg 0.0086 mg 0.013 mg 0.013 mg 0.019 mg 0.039 mg 0.054 mg 0.12 mg 0.23 mg 1.0 mg 1.4 mg 2.4 mg 5.0 mg |
| KC1001 Weight | CP-MA-0008 / Mettler MC210 S Sartorius LA32 00D Mettler XP260 03L Mettler KA30-3 | (0 to 200) g 200 g to 1 kg (1 to 26) kg (26 to 30) kg | 0.17 mg 2.0 mg 17 mg 72 mg |
| KC1002 Balance (On-site Calibrati on Inclu ded) | CP-MA-0007 /Mettler MC21 0S Sartorius L A3200D Mettle r XP26003L Me ttler KA30-3 | (0 to 5) g (5 to 20) g (20 to 50) g (50 to 100) g (100 to 210) g 210 g to 1 kg (1 to 6) kg (6 to 15) kg (15 to 26) kg (26 to 32) kg | 0.15 mg 0.24 mg 0.29 mg 0.46 mg 1.1 mg 5.0 mg 86 mg 0.22 g 0.37 g 0.46 g |
| KC1004 Platform Scale (On-site Calibrati on Inclu ded) | CP-MA-0007 /Mettler KC500 -1 | (32 to 100) kg (100 to 300) kg (300 to 600) kg | 5.2 g 11 g 17 g |
| KC1004 Crane Sc ale (On-site Calibrati on Inclu ded) | CP-MA-0013 /N.A. F1 | (0 to 600) kg(Resolution : 0.1 Kg) (0 to 600) kg(Resolution : 0.2 Kg) (0 to 1000) kg(Resolution : 0.2 Kg) (0 to 1000) kg(Resolution : 0.5 Kg) | 0.2 kg 0.4 kg 0.4 kg 1.0 kg |
| KC2001 Proving r ing | CP-FO-0002 /Morehouse Se ries 1000 | (0.98 to 440) kN | Tension 0.12 % Compression 0.12 % of full scale |
| KC2002 Load cell | CP-FO-0002 /Morehouse Se ries 1000 | (0.98 to 440) kN | Tension 0.16 % Compression 0.16 % of full scale |
| KC2005 Material testing (on-site c alibration included) | CP-FO-0003 /HBM Ormond/ S9, BUL-FF39- machine(OS-**-90-251 | (0.44 to 440) kN | Tension 0.16 % Compression 0.16 % of full scale |
| KC3001 Standard | CP-VO-0001 /Sartorius LA3 | (0 to 50) mL (50 to 100) mL | 0.12 mL 0.17 mL |

| | | | |
|---|-------------------------------------|---|--|
| Tank | 200D Mettler K C500-1 | (100 to 250) mL (250 to 500) mL (500 to 1000) mL (1000 to 2000) mL 10 L 20 L 50 L | 0.4 mL 0.8 mL 1.5 mL 1.6 mL 1.4 mL 2.6 mL 3.5 mL |
| KC4001 Torque Wrench | CP-TQ-0001 /Norbar 50031. ETS | (0.015 to 1000) N m | 2.0 % |
| KC4002 Torque S crewdriv er | CP-TQ-0001 /Norbar 50027. ETS | (0.015 to 50) N m | 2.0 % |
| KC4003 Torque S ensor/Ga uge/ Tra nsducer | CP-TQ-0002 /Norbar 21515 | (0.015 to 5424) N m | 0.1 % |
| KC4004 Torque C alibrator | CP-TQ-0002 /Norbar 21515 | (0.015 to 5424) N m | 0.1 % |
| KC4005 Torque Multiplier | CP-TQ-0003 /Norbar 50075. ETS | (1.0 to 4000) N m | 1.0 % |

Pressure/Vacuum

| Calibrati on items | Calibration met hod /Working Stan dard | Ranges | Smallest uncertainty |
|---|--|--|--|
| KD1001 Gas lubri cated pi ston pre ssure ga uge | CP-PR-0010 CP-PR-0360 /RUSKA 2465 R USKA 2400 | 1.38 kPa to 6.89 MPa 41.4 kPa to 83.7 MPa | 5.4 E-05 1.1 E-04 |
| KD1002 Oil lubri cated pist on press ure gaug e | CP-PR-0011 CP-PR-0360 /RUSKA 2400 R USKA 2485 | 41.4 kPa to 83.7 MPa 68.9 kPa to 414 MPa | 1.1 E-04 1.5 E-04 |
| KD1003 Mercury manomet er (on-site calibrati on include d) | CP-PR-0100 CP-PR-0360 /RUSKA 2465, Schwien 1025L X110 | 0 Pa to 373 kPa(RUSKA 2465) 0 Pa to 373 kPa(SCHWIEN 1025LX110) | 0.95 Pa + 2.2 E-05 1.7 Pa + 2.2 E-05 |
| KD1004 Pressure gauge (on-site calibratio n include d) | CP-PR-0360 CP-PR-0361 /SCHWIEN 102 5LX110 RUSKA 2465 RUSKA 2 400 RUSKA 24 85 | 0 Pa to 373 kPa 1.38 kPa to 6.89 MPa 41.4 kPa to 83.7 MPa 68.9 kPa to 414 MPa | 1.7 Pa + 2.2 E-5 5.4 E-05 1.1 E-04 1.5 E-04 |
| KD1005 Pressure transduc er (on-site calibratio n include d) | CP-PR-0360 CP-PR-0361 /SCHWIEN 102 5LX110 RUSKA 2465 RUSKA 2 400 RUSKA 24 85 | 0 Pa to 373 kPa 1.38 kPa to 6.89 MPa 41.4 kPa to 83.7 MPa 68.9 kPa to 414 MPa | 1.7 Pa + 2.2 E-5 5.4 E-05 1.1 E-04 1.5 E-04 |

| | | | |
|---|--|---|--|
| n include d) | 85 | | |
| KD1006 Air data test syst em (on-site calibratio n include d) | CP-PR-2120 /RUSKA 2465 S CHWIEN 1025L X110 RUSKA 7 750i | Altitude -914.40 m to 0.00 m (-3 kft to 0 kft) 0.00 m to 9144.00 m (0 kft to 30 kft) 9144.00 m to 15240.00 m (30 kft to 50 kft) 15240.0 m to 25384.0 m (50 kft to 80 kft) 24384.0 m to 30480.0 m (80 kft to 100 kft) Air Speed 12.861 m/s to 33.439 m/s (25 Knots to 65 Knots) 33.439 m/s to 51.444 m/s (65 Knots to 100 Knots) 51.444 m/s to 128.6111 m/s (100 Knots to 250 Knots) 128.6111 m/s to 180.0556 m/s (250 Knots to 350 Knots) 180.0556 m/s to 514.4444 m/s (350 Knots to 1000 Knots) | 0.37 m(1.2 ft) 0.43 m(1.4 ft) 0.67 m(2.2 ft) 2.4 m(7.8 ft) 6.1 m(20 ft) 0.062 m/s(0.12 Knots) 0.024 m/s(0.046 Knots) 0.016 m/s(0.030 Knots) 0.0072 m/s(0.014 Knots) 0.0057 m/s(0.011 knots) |
| KD1007 Differenti al pressu re gauge (on-site calibratio n include d) | CP-PR-0360 CP-PR-0361 /SCHWIEN 102 5LX110 RUSKA 2465 | 0 Pa to 373 kPa 1.38 kPa to 6.89 MPa | 1.7 Pa + 2.2 E-0.5 5.4 E-5 |
| KD2001 Spinning rotor vis cosity v acuum g auge | (CP-VA-0001) /MKS SRG-2 M KS 390HA-000 1 (BAZERS PSK 100) | (5.00 E-05 to 1.33 E-04) Pa (>1.33 E-04 to 1.33 E-03) Pa (>1.33 E-03 to 1.33 E+01) Pa | 32 % 12 % 4.0 % |
| KD2002 Ionizatio n vacuu m gauge (on-site calibratio n include d) | CP-VA-0001 /MKS SRG-2 M KS 390HA-000 1 (BAZERS PSK 100) | (5.00 E-05 to 1.33 E-04) Pa (>1.33 E-04 to 1.33 E-03) Pa (>1.33 E-03 to 1.33 E-01) Pa | 33 % 13 % 7.2 % |
| KD2003 Capacita nce diap hragm v acuum g auge (on-site calibratio n include d) | CP-VA-0001 CP-VA-0140 /MKS 390HA-0 1000 MKS 390 HA-00001 (MK S PVS-6B-HA) | (1.33 E-01 to 1.33 E+00) Pa (> 1.33 E+00 to 1.33 E+03) Pa (> 1.33 E+03 to 1.33 E+05) Pa | 2.7 % 1.5 % 0.32 % |
| KD2004 Compres sion vau um gaug e (on-site | CP-VA-0001 /MKS 390HA-0 1000 MKS 390 HA-00001 (MK S PVS-6B-HA) | (1.33 E-01 to 1.33 E+00) Pa (> 1.33 E+00 to 1.33 E+03) Pa | 3.5 % 1.9 % |

| | | | |
|---|--|---|--------------------------|
| calibration included) | | | |
| KD2005 Thermocouple vacuum gauge (on-site calibration included) | CP-VA-0001 /MKS 390HA-01000 MKS 390 HA-00001 (MKS PVS-6B-HA) | (1.33 E-01 to 1.33 E+00) Pa (> 1.33 E+00 to 1.33 E+03) Pa (> 1.33 E+03 to 1.33 E+05) Pa | 5.0 % 4.3 % 3.6 % |
| KD2006 Pirani vacuum gauge Digital pressure gauge Mechanical Vacuum gauge (on-site calibration included) | CP-VA-0001 /MKS 390HA-01000 MKS 390 HA-00001 (MKS PVS-6B-HA) | (1.33 E-01 to 1.33 E+00) Pa (> 1.33 E+00 to 1.33 E+03) Pa (> 1.33 E+03 to 1.33 E+05) Pa | 3.5 % 1.9 % 0.34 % |

Temperature/Humidity

| Calibration items | Calibration method /Working Standard | Ranges | Smallest uncertainty |
|---|---|---|--|
| KE1001 Liquid-in-glass Thermometer | Thermometer calibration procedure (CP-TE-0010) /HART 5626 | (-50 to 250) °C | 0.09 °C |
| KE1002 Platinum Resistance Thermometer | Thermometer calibration procedure (CP-TE-0010) /HART 5626/HART5638 | (-50 to 250) °C (> 250 to 600) °C (>600 to 1000) °C | 0.033 °C 0.057 °C 0.83 °C |
| KE1004 Thermocouple | Thermocouple calibration procedure (CP-TE-0010) /HART 5626/ HART 5638 | (-50 to 600) °C (>600 to 1000) °C | 0.50 °C 1.0 °C |
| KE1005 Thermocouple Thermometer (in Lab) | Thermometer calibration procedure (CP-TE-0010) /HART 5626/ HART 5638 | (-50 to 600) °C (>600 to 1000) °C | 0.50 °C 1.0 °C |
| KE1006 Temperature Indicator | Temperature indicator calibration procedure (CP-TE-0004) /EDC 522 | J TYPE (-50 to 1000) °C K TYPE (-50 to 1000) °C T TYPE (-50 to 400) °C E TYPE (-50 to 1000) °C | 0.49 °C 0.49 °C 0.38 °C 0.49 °C |

| | | | |
|--|--|---|--|
| | | R TYPE (0 to 1000) °C S TYPE (0 to 1000) °C RTD (0 to 600) °C | 0.38 °C 0.38 °C 0.1 °C |
| KE1008 Thermocouples Simulator | Thermocouples simulator calibration procedure (CP-TE-0004) /EDC 522 | J TYPE (-50 to 1000) °C K TYPE (-50 to 1000) °C T TYPE (-50 to 400) °C E TYPE (-50 to 1000) °C R TYPE (0 to 1000) °C S TYPE (0 to 1000) °C RTD (-50 to 600) °C | 0.49 °C 0.49 °C 0.38 °C 0.49 °C 0.38 °C 0.38 °C 0.1 °C |
| KE1009 Temperature Calibrator | Temperature calibrator calibration procedure (CP-TE-0010) /HART 5626 HART 5638 | (-50 to 600) °C (>600 to 1000) °C | 0.50 °C 1.0 °C |
| KE1010 Temperature Chamber, Temperature Furnace, Oven, Dry Oven (on-site calibration) | CP-TE-0005 /FLUKE 724 | (-50 to 1000) °C | 1.2 °C |
| KE1011 Thermocouple Thermometer, Temperature Indicator (Temperature Chamber, Temperature Furnace, Oven, Dry Oven) (on-site calibration) | Temperature on-site calibration procedure (CP-TE-0005) /FLUKE 724 | (-50 to 1000) °C (-50 to 1000) °C | 1.2 °C 0.9 °C |
| KE2001 Hygrometer | Hygrometer calibration procedure (CP-HU-0001) /THUNDER 250 OST | (10 to 90) %RH (5 to 60) °C | 1.0 %RH 0.2 °C |
| KE2003 Dew point hygrometer | Dew point hygrometer calibration procedure | (-60 to 0) °C | 0.9 °C |

| | | | |
|--|-------------------|---|--------------------------|
| Dew Point Hygrometer | Humidity Recorder | ion procedure (CP-HU-0002) /RH SYSTEMS 373L | |
| KE2004 Temperature & Humidity Recorder | | (CP-HU-0001) /THUNDER 250 OST (5 to 60) °C | (10 to 90) %RH 0.7 °C |

Electricity

| Calibration items | Calibration method /Working Standard | Ranges | Smallest uncertainty |
|--|--------------------------------------|--|--|
| KF1001 DCV Source | CP-SD-V001, CP-SD-0001, CP-MT-0001 | 1 mV 10 mV 0.1 V 1.0 V | 1.1 mV/V 0.13 mV/V 7.8 μV/V 1.3 μV/V |
| KF1001 DCV Source | /DATRON 4910 | 1.018 V | 1.3 μV/V |
| KF1001 DCV Meter | , FLUKE 5700A, DATRON 1281 | 10 V 100 V 1000 V | 1.2 μV/V 4.5 μV/V 5.6 μV/V |
| KF1001 | CP-SD-0004 / FLUKE 5700A | < 100 mV 0.1 V to 1.0 V 1.0 V to 10 V 10 V to 100 V 100 V to 1000V | 32 μV/V 17 μV/V 17 μV/V 21 μV/V 23 μV/V |
| KF1002 DC Current Source | CP-SD-0001, CP-MT-0001 | 100 μA 1 mA 10 mA 100 mA | 71 μA/A 44 μA/A 46 μA/A 60 μA/A |
| KF1002 DC Current Meter | /FLUKE 5700A, DATRON 1281 | 1 A | 67 μA/A |
| KF1002 | CP-SD-0004 /FLUKE 5700A | < 100 μA 0.1 mA to 1.0 mA 1.0 mA to 10 mA 10 mA to 100 mA 100 mA to 1000 mA | 94 μA/A 52 μA/A 54 μA/A 67 μA/A 96 μA/A |
| KF1011 AC/DC Thermal Transfer Standard | CP-SD-0002 /FLUKE 792A | 2 mV(10 Hz) 2 mV(20 Hz) 2 mV(40 Hz) 2 mV(100 Hz) 2 mV(1 kHz) 2 mV(10 kHz) 2 mV(20 kHz) 2 mV(50 kHz) 2 mV(100 kHz) 2 mV(300 kHz) 2 mV(500 kHz) 2 mV(1 MHz) | 2.8 mV/V 2.1 mV/V 2.0 mV/V 2.1 mV/V 1.9 mV/V 2.0 mV/V 2.2 mV/V 1.9 mV/V 2.5 mV/V 3.2 mV/V 5.6 mV/V 6.5 mV/V |
| KF1011 | / | 10 mV(10 Hz) 10 mV(20 Hz) 10 mV(40 Hz) 10 mV(100 Hz) 10 mV(1 kHz) 10 mV(10 kHz) 10 mV(20 kHz) 10 mV(50 kHz) 10 mV(100 kHz) 10 mV(300 kHz) | 0.39 mV/V 0.39 mV/V 0.34 mV/V 0.36 mV/V 0.40 mV/V 0.36 mV/V 0.36 mV/V 0.47 mV/V 0.67 mV/V 0.85 mV/V |

| | | | |
|--------|---|--|---|
| | | 10 mV(500 kHz) 10 mV(1 MHz) | 1.6 mV/V 1.8 mV/V |
| KF1011 | / | 20 mV(10 Hz) 20 mV(20 Hz) 20 mV(40 Hz) 20 mV(100 Hz) 20 mV(1 kHz) 20 mV(10 kHz) 20 mV(20 kHz) 20 mV(50 kHz) 20 mV(100 kHz) 20 mV(300 kHz) 20 mV(500 kHz) 20 mV(1 MHz) | 0.34 mV/V 0.22 mV/V 0.18 mV/V 0.19 mV/V 0.18 mV/V 0.18 mV/V 0.18 mV/V 0.28 mV/V 0.58 mV/V 0.83 mV/V 1.4 mV/V 1.6 mV/V |
| KF1011 | / | 100 mV(10 Hz) 100 mV(20 Hz) 100 mV(40 Hz) 100 mV(100 Hz) 100 mV(1 kHz) 100 mV(10 kHz) 100 mV(20 kHz) 100 mV(50 kHz) 100 mV(100 kHz) 100 mV(300 kHz) 100 mV(500 kHz) 100 mV(1 MHz) | 0.25 mV/V 0.13 mV/V 0.060 mV/V 0.070 mV/V 0.067 mV/V 0.065 mV/V 0.065 mV/V 0.13 mV/V 0.19 mV/V 0.54 mV/V 0.72 mV/V 0.74 mV/V |
| KF1011 | / | 200 mV(10 Hz) 200 mV(20 Hz) 200 mV(40 Hz) 200 mV(100 Hz) 200 mV(1 kHz) 200 mV(10 kHz) 200 mV(20 kHz) 200 mV(50 kHz) 200 mV(100 kHz) 200 mV(300 kHz) 200 mV(500 kHz) 200 mV(1 MHz) | 0.26 mV/V 0.11 mV/V 0.064 mV/V 0.065 mV/V 0.064 mV/V 0.061 mV/V 0.065 mV/V 0.13 mV/V 0.19 mV/V 0.54 mV/V 0.71 mV/V 0.73 mV/V |
| KF1011 | / | 600 mV(10 Hz) 600 mV(20 Hz) 600 mV(40 Hz) 600 mV(100 Hz) 600 mV(1 kHz) 600 mV(10 kHz) 600 mV(20 kHz) 600 mV(50 kHz) 600 mV(100 kHz) 600 mV(300 kHz) 600 mV(500 kHz) 600 mV(1 MHz) | 0.24 mV/V 0.11 mV/V 0.043 mV/V 0.032 mV/V 0.033 mV/V 0.032 mV/V 0.030 mV/V 0.061 mV/V 0.082 mV/V 0.18 mV/V 0.51 mV/V 0.65 mV/V |
| KF1011 | / | 1 V (10 Hz) 1 V(20 Hz) 1 V(40 Hz) 1 V(100 Hz) 1 V(1 kHz) 1 V(10 kHz) 1 V(20 kHz) 1 V(50 kHz) | 0.25 mV/V 0.078 mV/V 0.037 mV/V 0.019 mV/V 0.020 mV/V 0.020 mV/V 0.020 mV/V 0.055 mV/V |

| | | | |
|--------|---|--|--|
| | | 1 V(100 kHz) 1 V(300 kHz) 1 V(500 kHz) 1 V(1 MHz) | 0.068 mV/V 0.16 mV/V 0.51 mV/V 0.53 mV/V |
| KF1011 | / | 2 V(10 Hz) 2 V(20 Hz) 2 V(40 Hz) 2 V(100 Hz) 2 V(1 kHz) 2 V(10 kHz) 2 V(20 kHz) 2 V(50 kHz) 2 V(100 kHz) 2 V(300 kHz) 2 V(500 kHz) 2 V(1 MHz) | 0.24 mV/V 0.084 mV/V 0.043 mV/V 0.034 mV/V 0.032 mV/V 0.032 mV/V 0.032 mV/V 0.063 mV/V 0.074 mV/V 0.17 mV/V 0.53 mV/V 0.56 mV/V |
| KF1011 | / | 6 V(10 Hz) 6 V(20 Hz) 6 V(40 Hz) 6 V(100 Hz) 6 V(1 kHz) 6 V(10 kHz) 6 V(20 kHz) 6 V(50 kHz) 6 V(100 kHz) 6 V(300 kHz) 6 V(500 kHz) 6 V(1 MHz) | 0.25 mV/V 0.083 mV/V 0.044 mV/V 0.030 mV/V 0.032 mV/V 0.034 mV/V 0.032 mV/V 0.055 mV/V 0.074 mV/V 0.17 mV/V 0.52 mV/V 0.57 mV/V |
| KF1011 | / | 10 V(10 Hz) 10 V(20 Hz) 10 V(40 Hz) 10 V(100 Hz) 10 V(1 kHz) 10 V(10 kHz) 10 V(20 kHz) 10 V(50 kHz) 10 V(100 kHz) 10 V(300 kHz) 10 V(500 kHz) 10 V(1 MHz) | 0.23 mV/V 0.081 mV/V 0.037 mV/V 0.025 mV/V 0.025 mV/V 0.024 mV/V 0.025 mV/V 0.054 mV/V 0.069 mV/V 0.16 mV/V 0.51 mV/V 0.57 mV/V |
| KF1011 | / | 20 V(10 Hz) 20 V(20 Hz) 20 V(40 Hz) 20 V(100 Hz) 20 V(1 kHz) 20 V(10 kHz) 20 V(20 kHz) 20 V(50 kHz) 20 V(100 kHz) 20 V(300 kHz) 20 V(500 kHz) 20 V(1 MHz) | 0.24 mV/V 0.085 mV/V 0.042 mV/V 0.034 mV/V 0.033 mV/V 0.032 mV/V 0.034 mV/V 0.073 mV/V 0.096 mV/V 0.20 mV/V 0.51 mV/V 0.57 mV/V |
| KF1011 | / | 60 V(10 Hz) 60 V(20 Hz) 60 V(40 Hz) 60 V(100 Hz) 60 V(1 kHz) 60 V(10 kHz) | 0.25 mV/V 0.086 mV/V 0.053 mV/V 0.052 mV/V 0.050 mV/V 0.051 mV/V |

| | | | |
|--------|---|--|--|
| | | 60 V(20 kHz) 60 V(50 kHz) 60 V(100 kHz) 60 V(300 kHz) | 0.055 mV/V 0.085 mV/V 0.097 mV/V 0.17 mV/V |
| KF1011 | / | 100 V(10 Hz) 100 V(20 Hz) 100 V(40 Hz) 100 V(100 Hz) 100 V(1 kHz) 100 V(10 kHz) 100 V(20 kHz) 100 V(50 kHz) 100 V(100 kHz) | 0.26 mV/V 0.080 mV/V 0.043 mV/V 0.040 mV/V 0.038 mV/V 0.038 mV/V 0.038 mV/V 0.085 mV/V 0.092 mV/V |
| KF1011 | / | 200 V(10 Hz) 200 V(20 Hz) 200 V(40 Hz) 200 V(100 Hz) 200 V(1 kHz) 200 V(10 kHz) 200 V(20 kHz) 200 V(50 kHz) 200 V(100 kHz) | 0.27 mV/V 0.12 mV/V 0.060 mV/V 0.051 mV/V 0.049 mV/V 0.052 mV/V 0.055 mV/V 0.084 mV/V 0.11 mV/V |
| KF1011 | / | 1000 V(40 Hz) 1000 V(100 Hz) 1000 V(1 kHz) 1000 V(10 kHz) 1000 V(20 kHz) | 0.037 mV/V 0.036 mV/V 0.037 mV/V 0.038 mV/V 0.042 mV/V |
| KF1011 | CP-SD-0001, ACV Source CP-MT-0001 ACV Meter /FLUKE 5700A, DATRON 1281 | 0.1 V 1.0 V 10 V 100 V 1000 V (1 kHz) 0.1 V 1.0 V 10 V 100 V 1000 V (60 Hz) | 0.25 mV/V 0.092 mV/V 0.092 mV/V 0.12 mV/V 0.13 mV/V 0.25 mV/V 0.092 mV/V 0.092 mV/V 0.12 mV/V 0.13 mV/V |
| KF1011 | CP-SD-0004 /FLUKE 5700A | < 100 mV 0.1 V to 1.0 V 1.0 V to 10 V 10 V to 100 V 100 V to 1000 V (1 kHz) | 0.26 mV/V 0.11 mV/V 0.11 mV/V 0.13 mV/V 0.15 mV/V |
| KF1012 | CP-SD-0001, C AC Current Source P-MT-0001 AC Current Meter /FLUKE 5700A, DATRON-1281 | 100 µA 1 mA 10 mA 100 mA 1 A (1 kHz) 100 µA 1 mA 10 mA 100 mA 1 A (60 Hz) | 0.45 mA/A 0.20 mA/A 0.20 mA/A 0.22 mA/A 0.88 mA/A 0.45 mA/A 0.20 mA/A 0.20 mA/A 0.22 mA/A 0.88 mA/A |
| KF1012 | | | |

| | | | |
|---|--|---|---|
| | CP-SD-0004 / FLUKE 5700A | ACA at 1 kHz < 100 μ A 0.1 mA to 1.0 mA 1.0 mA to 10 mA 10 mA to 100 mA 100 mA to 1000 mA | 0.56 mA/A 0.23 mA/A 0.23 mA/A 0.25 mA/A 0.92 mA/A |
| KF3001 Resistor | Resistance calibration procedure(CP-SD-R0001) / ESI SR104, L&N4210B | 0.001 Ω 0.01 Ω 0.1 Ω 1 Ω 10 Ω 100 Ω 1 k Ω 10 k Ω 100 k Ω 1 M Ω | 68 $\mu\Omega/\Omega$ 50 $\mu\Omega/\Omega$ 38 $\mu\Omega/\Omega$ 2.4 $\mu\Omega/\Omega$ 6.4 $\mu\Omega/\Omega$ 8.8 $\mu\Omega/\Omega$ 6.0 $\mu\Omega/\Omega$ 1.2 $\mu\Omega/\Omega$ 12 $\mu\Omega/\Omega$ 17 $\mu\Omega/\Omega$ |
| KF3001 Ohmmeter Multimeter | Calibration procedure of calibrator(CP-SD-0001), Calibration procedure of DVM(CP-MT-0001) / FLUKE-5700A DATRON-1281 | 1 Ω 10 Ω 100 Ω 1 k Ω 10 k Ω 100 k Ω 1 M Ω 10 M Ω | 0.14 m Ω/Ω 41 $\mu\Omega/\Omega$ 27 $\mu\Omega/\Omega$ 20 $\mu\Omega/\Omega$ 18 $\mu\Omega/\Omega$ 30 $\mu\Omega/\Omega$ 31 $\mu\Omega/\Omega$ 58 $\mu\Omega/\Omega$ |
| KF3001 Resistor | (CP-SD-0004) / FLUKE-5700A | 1 Ω to 10 Ω 10 Ω to 100 Ω 100 Ω to 1 k Ω 1 k Ω to 10 k Ω 10 K Ω to 100 k Ω 100 k Ω to 1 M Ω 1 M Ω to 10 M Ω | 0.15 m Ω/Ω 46 $\mu\Omega/\Omega$ 50 $\mu\Omega/\Omega$ 45 $\mu\Omega/\Omega$ 49 $\mu\Omega/\Omega$ 34 $\mu\Omega/\Omega$ 0.28 m Ω/Ω |
| KF3002 Inductor Inductance Meter /Bridge | Inductance calibration procedure (CP-SD-L001) / GR-1689M | 100 μ H 1 mH 10 mH 100 mH 1 H 10 H (100 Hz) 100 μ H 1 mH 10 mH 100 mH 1 H 10 H (1 kHz) | 3.0 mH/H 0.6 mH/H 0.6 mH/H 0.6 mH/H 0.6 mH/H 0.6 mH/H (100 Hz) 3.0 mH/H 0.5 mH/H 0.6 mH/H 0.6 mH/H 2.0 mH/H 5.0 mH/H (1 kHz) |
| KF3003 Capacitor Capacitor Meter/ Bridge | Capacitance calibration procedure(CP-SD-C001) / AH-2500A | 1 pF 10 pF 100 pF 1000 pF 0.01 μ F 0.1 μ F 1 μ F (1 kHz) | 7.0 μ F/F 6.0 μ F/F 6.0 μ F/F 13 μ F/F 0.15 mF/F 0.15 mF/F 0.15 mF/F (1 kHz) |

Electromagnetics

| Calibration items | Calibration method /Working Standard | Ranges | Smallest uncertainty |
|-------------------|---|--------|----------------------|
|-------------------|---|--------|----------------------|

| | | | |
|--------|---------------|-------------------------------|-------------------------|
| KC1002 | Microwave Sca | Frequency: 45 Mhz to 26.5 GHz | Reflection Coefficient: |
|--------|---------------|-------------------------------|-------------------------|

| | | | |
|--|--|--|--|
| RG1002 Attenuator, Terminator, Open, Short, Mismatch, Microwave Device, Vector Network Analyzer | Microwave Scattering Parameters and Impedance Calibration Procedure (CP-SD-1702) /HP 8510C | Frequency: 45 MHz to 26.5 GHz 1. Reflection Coefficient: 0 to 1 2. Transmission Coefficient: 10 dB to -60 dB Reflection Coefficient: 45 MHz to 2 GHz 3.5 mm Type N Type 7 mm Type > 2 GHz to 8 GHz 3.5 mm Type N Type 7 mm Type > 8 GHz to 20 GHz 3.5 mm Type > 8 GHz to 18 GHz N Type 7 mm Type > 20 GHz to 26.5 GHz 3.5 mm Type Transmission Coefficient: 45 MHz to 2 GHz 3.5 mm Type N Type 7 mm Type > 2 GHz to 8 GHz 3.5 mm Type N Type 7 mm Type > 8 GHz to 20 GHz 3.5 mm Type > 8 GHz to 18 GHz N Type 7 mm Type > 20 GHz to 26.5 GHz 3.5 mm Type | Reflection Coefficient: 0.0062 (1.17°) 0.0062 (0.93°) 0.013 (1.63°) 0.0085 (3.30°) 0.011 (3.17°) 0.013 (3.50°) 0.0086 (5.78°) 0.011 (5.53°) 0.013 (5.98°) 0.0087 (7.05°) Transmission Coefficient: 0.054 dB (0.21°) 0.051 dB (0.21°) 0.059 dB (0.25°) 0.064 dB (1.80°) 0.061 dB (1.79°) 0.061 dB (1.78°) 0.094 dB (4.29°) 0.092 dB (3.91°) 0.085 dB (3.84°) 0.11 dB (5.64°) |
|--|--|--|--|

Flow

| Calibration items | Calibration method /Working Standard | Ranges | Smallest uncertainty |
|--------------------------|--|-----------------------------------|----------------------|
| KH1001 Gas flow meter | (CP-FL-0004) /Brooks Bell Prover 1092A | (3.6 to 300) dm ³ /min | 0.45 % |
| KH1002 Flow meter | (CP-FL-0002) /EG&G OT4002 T1S | (0.4 to 250) GPM | 0.16 % |
| KH1003 Flow meter | (CP-FL-0001) /EG&G MT501T 1A | (0.05 to 50) GPM | 0.23 % |

Chemical

| Calibration items | Calibration method /Working Standard | Ranges | Smallest uncertainty |
|-------------------|--------------------------------------|--------|----------------------|
|-------------------|--------------------------------------|--------|----------------------|

| | | | |
|----------------------|--|--------------------------------|--------|
| | dard | | |
| KI3000 Hydrometer | (CP-CM-0001) /Anton Pear D MA 58 | 0.7990 to 1.5160 (@ 20 °C) | 0.0004 |

Time And Frequency

| Calibration items | Calibration method /Working Standard | Ranges | Smallest uncertainty |
|--|--------------------------------------|--------------------------|----------------------|
| KJ0100 Timer | CP-TM-0001 /HP5071A | (0 to 900) s | 5.0 E-02 s |
| KJ0200 Frequency Standard Standard Frequency Standard etc. | CP-FN-0035 /HP5071A | 1 MHz 5 MHz 10 MHz | 8.2 E-11 |
| KJ0200 Frequency counter | CP-FN-F002 /HP5071A | 100 KHz to 1.3 GHz | 6.4 E-09 |
| KJ0300 Tachometer | CP-TA-0001 /TICO 8730 | (20 to 15000) rpm | 2 rpm |

Approval Signatory

| Approval Signatory | Scope |
|--------------------|--|
| HO, Chieh | KA1001,KA1003,KA1009,KA1019,KA2003,KA2004,KA2005,KA2007,KA2008,KA2010,KA2012,KA3001,KA3002,KA3003,KA3004,KA3005,KA3008,KA3010,KA3015,KA3016,KA3019,KA4002,KA4003 |
| LIU, Yu-San | KG1002,KJ0100,KJ0200,KJ0300 |
| SHAI, Shaw-Kang | KF1001,KF1002,KF1011,KF1011,KF1012,KF3001,KF3002,KF3003 |
| LEE, Kuo-Chi | KB1001,KB1002,KB2004,KC4001,KC4002,KC4003,KC4004,KC4005,KE1001,KE1002,KE1004,KE1005,KE1006,KE1008,KE1009,KE1010,KE1011,KE2001,KE2003,KE2004 |
| LEE, Yuh-Tay | KJ0100,KJ0200,KJ0300 |
| YANG, Chih-Ying | KF1001,KF1002,KF1011,KF1011,KF1012,KF3001,KF3002,KF3003 |
| LO, Yun-Shao | KC4001,KC4002,KC4003,KC4004,KC4005,KH1001,KH1002,KH1003 |
| CHUANG, Chao-Chang | KC1001,KC1002,KC1004,KC2001,KC2002,KC2005,KC3001,KE2001,KE2003,KE2004,KH1001,KH1002,KH1003,KI3000 |
| YEH, Dong-Ming | KG1002,KJ0100,KJ0200,KJ0300 |
| TSAI, Yuan-Hsing | KA1001,KA1003,KA1009,KA1019,KA2005,KA2007,KC1001,KC1002,KC1004,KC3001 |
| CHANG, Jong-Ping | KA2003,KA2004,KA2008,KA2010,KA2012 |
| HUANG, Chao-Yu | KB1001,KB1002,KB2004,KC1001,KC1002,KC1004,KC3001,KD1001,KD1002,KD1003,KD1004,KD1005,KD1006,KD1007,KD2001,KD2002,KD2003,KD2004,KD2005,KD2006 |

若全部項目皆為保留，是否需委員審查 是 否

認證經理簽名：林宜臻 日期(Date)：2014/05/19

歷史資料